

# Time & Frequency Bulletin

No. 383 October 1989

(NISTIR 89-3910-10)

National  
Institute of  
Standards &  
Technology





**NIST TIME & FREQUENCY BULLETIN**  
(Supersedes No. 382 September 1989)

NO. 383

OCTOBER 1989

1. GENERAL BACKGROUND INFORMATION . . . . .	1
2. TIME SCALE INFORMATION . . . . .	1
International Timing Center comparisons via GPS common-view . . . . .	2
3. UT1 CORRECTIONS AND LEAP SECOND ADJUSTMENTS . . . . .	2
4. PHASE DEVIATIONS FOR WWVB AND LORAN-C . . . . .	3
5. GOES TIME CODE INFORMATION . . . . .	4
6. BROADCAST OUTAGES OVER FIVE MINUTES AND WWVB PHASE PERTURBATIONS . . . . .	4
7. NOTES ON NIST TIME SCALES AND PRIMARY STANDARDS . . . . .	5
8. SPECIAL ANNOUNCEMENTS . . . . .	7

This bulletin is published monthly. Address correspondence to:

Gwen E. Bennett, Editor  
Time and Frequency Division  
National Institute of Standards & Technology  
325 Broadway  
Boulder, CO 80303  
(303) 497-3295

NOTE TO SUBSCRIBERS: Please include your address label (or a copy)  
with any correspondence regarding this bulletin.

1. GENERAL BACKGROUND INFORMATION

ABBREVIATIONS AND ACRONYMS USED IN THIS BULLETIN

APL	-	John Hopkins University Applied Physics Laboratory	
BIH	-	International Time Bureau, France	
CCIR	-	International Radio Consultative Committee	
CRL	-	Communications Research Laboratories, Japan	
Cs	-	Cesium standard	
CSIRO	-	Commonwealth Scientific and Industrial Research Organization, Australia	
GOES	-	Geostationary Operational Environmental Satellite	
GPS	-	Global Positioning System	
IEN	-	National Institute of Electronics, Italy	
INPL	-	National Physical Laboratory, Israel	
LORAN	-	Long Range Navigation	
MC	-	Master Clock	
MJD	-	Modified Julian Date	
NIST	-	National Institute of Standards & Technology	
NPL	-	National Physical Laboratory, England	
NRC	-	National Research Council, Canada	
NOAA	-	National Oceanic and Atmospheric Administration	
OP	-	Paris Observatory, France	
PTB	-	Physical Technical Federal Laboratory, Germany	
SI	-	International System of Units	ns - nanosecond
SV	-	Space vehicle	μs - microsecond
TA	-	Atomic Time	ms - millisecond
TAI	-	International Atomic Time	s - second
TAO	-	Tokyo Astronomical Observatory, Japan	min - minute
TUG	-	Technical University of Graz, Austria	h - hour
USNO	-	United States Naval Observatory	d - day
UTC	-	Coordinated Universal Time	
VLF	-	very low frequency	
VSL	-	Van Swinden Laboratory, Netherlands	

2. TIME SCALE INFORMATION

The values listed below are based on data from the BIH, the USNO, and the NIST. The UTC - UTC(NIST) values are extrapolations since UTC is computed more than two months after the fact. The UTC(USNO,MC) - UTC(NIST) values are averaged measurements from NAVSTAR satellites 3,4,6, and 8 (see references on page 6).

0000 HOURS COORDINATED UNIVERSAL TIME

SEPTEMBER		UT1 - UTC(NIST)	UTC - UTC(NIST)	UTC(USNO,MC) - UTC(NIST)
1989	MJD	(± 5 ms)	(± 0.2 μs)	(± 0.04 μs)
7	47776	-456 ms	-0.5 μs	0.67 μs
14	47783	-463 ms	-0.5 μs	0.71 μs
21	47790	-475 ms	-0.4 μs	0.73 μs
28	47797	-484 ms	-0.4 μs	0.73 μs

INTERNATIONAL TIMING CENTER COMPARISONS VIA GPS COMMON-VIEW

The table below is a weighted average of the indicated GPS satellites used as transfer standards to measure the time difference of Timing Center (i) - UTC(NIST) by the simultaneous common-view approach (see references, page 6). The day-to-day variations of this technique are a few nanoseconds and the accuracy is about 10 ns. The time of the measurement is interpolated to 0000 UTC for the particular MJD ending in 9. These data are prepared for the BIPM for the computation on TAI and of UTC. All differential delays are 0 unless otherwise noted.

UTC(i) - UTC(NIST) (ns)		MJD			
UTC(i)	SV NUMBERS	47769	47779	47789	47799
UTC(APL) - UTC(NIST)	3, 12,13	1692*	1819	1804	5@
UTC(CRL) - UTC(NIST)	6, 12 14	720	709	717	735#
UTC(CSIRO) - UTC(NIST)	VIA CRL & TAO	22028	22268	22609	22885#
UTC(IEN) - UTC(NIST)	11,12,13	-693*	-606	-520	-473#
UTC(INPL) - UTC(NIST)	VIA OP	-131506*	-132738	-134028	-135287##
UTC(NPL) - UTC(NIST)	3, 11,12,13,14	677	804	968	1100##
UTC(NRC) - UTC(NIST)***	3,6,9,11,12,13,14	14430*	14624	14825	14994
UTC(OP) - UTC(NIST)	3, 11,12,13,14	516*	460	394	352##
UTC(PTB) - UTC(NIST)	11	-4367*	-4301	-4224	-4162#
UTC(TAO) - UTC(NIST)	6, 14	3616	3752	3901	4026@@
UTC(TUG) - UTC(NIST)	3, 11,12,13,14	-2101*	-2284	-2470	-2660#
UTC(USNO,MC) - UTC(NIST)	3,6,9,11,12,13,14	633*	686	722	725#
UTC(VSL) - UTC(NIST)	3, 11,12,13,14	-2015*	-2032	-2061	-2089#

\* These values have been updated from those printed in last month's Bulletin.

\*\*\* UTC(NRC) - UTC(NIST) has a differential delay of 41.2 ns; all other comparisons are computed using zero (0).

@ UTC(APL) experienced a time step of approximately -1792 ns on MJD 47799. The value given is computed from unfiltered data.

# These values for MJD 47799 were extrapolated forward from 47796.

## These values for MJD 47799 were extrapolated forward from 47795.

@@ These values for MJD 47799 were extrapolated forward from 47797.

3. UT1 CORRECTIONS AND LEAP SECOND ADJUSTMENTS

The master clock pulses used by the WWV, WWVH, WWVB, and GOES time code transmissions are referenced to the UTC(NIST) time scale. Occasionally, 1 second is added to the UTC time scale. This second is called a leap second. Its purpose is to keep the UTC time scale within ± 0.9 s of the UT1 astronomical time scale, which changes slightly due to variations in the rotation of the earth.

Positive leap seconds, beginning at 23 h 59 min 60 s UTC and ending at 0 h 0 min 0 s UTC, were inserted in the UTC timescale on 30 June 1972, 31 December 1972-1979, 30 June 1981-1983, 30 June 1985, and 31 December 1987. When future leap seconds are scheduled, advance notice will be provided in this bulletin.

The use of leap seconds ensures that UT1 - UTC will always be held within ± 0.9 s. The current value of UT1 - UTC is called the DUT1 correction. DUT1 corrections are broadcast by WWV, WWVH, WWVB, and GOES and are printed below. These corrections may be added to received UTC time signals in order to obtain UT1.

DUT1 = UT1 - UTC	= -0.3 s beginning 0000 UTC 06 April 1989
	= -0.4 s beginning 0000 UTC 08 June 1989
	= -0.5 s beginning 0000 UTC 21 September 1989

**SPECIAL ANNOUNCEMENT:** A positive leap second will be introduced at the end of December 1989. The sequence of dates of the UTC second markers will be:

- 1989 December 31, 23h 59m 59s
- 1989 December 31, 23h 59m 60s
- 1990 January 1, 0h 0m 0s

#### 4. PHASE DEVIATIONS FOR WWVB AND LORAN-C

WWVB - The values shown for WWVB are the time difference between the time markers of the UTC(NIST) time scale and the first positive-going zero voltage crossover measured at the transmitting antenna. The uncertainty of the individual measurements is  $\pm 0.5 \mu\text{s}$ . The values listed are for 1500 UTC.

LORAN-C - The values shown for Loran-C represent the daily accumulated phase shift (in microseconds). The phase shift is measured by comparing the output of a Loran receiver to the UTC(NIST) time scale for a period of 24 hours. If data were not recorded on a particular day, the symbol (-) is printed.

The stations monitored are Dana, Indiana (8970 M) and Fallon, Nevada (9940 M). The monitoring is done from the NIST laboratories in Boulder, Colorado.

DATE	MJD	UTC(NIST) - WWVB(60 kHz) ANTENNA PHASE (in $\mu\text{s}$ )	UTC(NIST) - LORAN PHASE (in $\mu\text{s}$ )	
			LORAN-C (DANA) (8970 M)	LORAN-C (FALLON) (9940 M)
09/01/89	47770	5.65	-0.24	-0.21
09/02/89	47771	5.62	-0.13	+0.05
09/03/89	47772	5.59	(-)	+0.00
09/04/89	47773	5.55	(-)	+0.14
09/05/89	47774	5.53	(-)	+0.09
09/06/89	47775	5.72	-0.06	-0.15
09/07/89	47776	5.69	(-)	-0.02
09/08/89	47777	5.61	(-)	-0.25
09/09/89	47778	5.61	(-)	-0.15
09/10/89	47779	5.62	-0.18	-0.10
09/11/89	47780	5.62	(-)	(-)
09/12/89	47781	5.60	-0.09	+0.02
09/13/89	47782	5.67	(-)	(-)
09/14/89	47783	5.70	+0.15	+0.38
09/15/89	47784	5.64	-0.16	-0.10
09/16/89	47785	5.60	-0.23	-0.25
09/17/89	47786	5.55	+0.30	+0.20
09/18/89	47787	5.50	(-)	-0.10
09/19/89	47788	5.70	+0.06	+0.06
09/20/89	47789	5.63	+0.06	-0.21
09/21/89	47790	5.58	-0.16	+0.01
09/22/89	47791	5.63	-0.16	-0.20
09/23/89	47792	5.62	-0.16	+0.01
09/24/89	47793	5.60	(-)	+0.46
09/25/89	47794	5.58	-0.11	-0.11
09/26/89	47795	5.60	+0.31	+0.26
09/27/89	47796	5.66	-0.04	-0.09
09/28/89	47797	5.65	+0.17	+0.26
09/29/89	47798	5.64	-0.34	-0.58
09/30/89	47799	5.62	+0.05	+0.13

## 5. GOES TIME CODE INFORMATION

### A. TIME CODE PERFORMANCE (1 - 30 September 1989)

GOES/East: Performance within normal limits during this period.

GOES/West: Performance within normal limits during this period.

B. SPECIAL REMINDER: Current satellite locations are 65° West longitude for GOES/East and 135° West longitude for GOES/West.

### C. GOES STATUS REPORTS

A brief message from the NIST giving current GOES time code status information is available from the U.S. Naval Observatory's Automated Data Service computer system in Washington, DC. The message may be accessed 24 hours per day without charge by using a variety of terminals operating at 300, 1200, or 2400 Baud and even parity. Two different sets of telephone access numbers are available: (1) for 300 or 1200 Baud and the Bell 103 standard use (202) 653-1079 (commercial), 653-1079 (FTS), or 294-1079 (Autovon); (2) for 1200 or 2400 Baud with either the CCITT V.22 standard or the Bell standard use (202) 653-1783 (commercial), 653-1783 (FTS), or 294-1783 (Autovon). To receive the GOES status message, use the following procedure:

1. Access the USNO computer database by dialing one of the appropriate telephone numbers above;
2. In response to the prompt for identification, type your name and the name of your organization, followed by a carriage return;
3. Type "@NBSGO" followed by a carriage return to receive the status message at your terminal;
4. Disconnect by typing Control-D.

## 6. BROADCAST OUTAGES OVER 5 MINUTES AND WWVB PHASE PERTURBATIONS

OUTAGES						PHASE PERTURBATIONS WWVB 60 kHz			
STATION	SEPTEMBER 1989	MJD	BEGAN (UTC)	ENDED (UTC)	FREQUENCY	SEPTEMBER 1989	MJD	BEGAN (UTC)	ENDED (UTC)
WWVB	NONE								
WWV	NONE								
WWVH	25	47794	0318:00	0411:00	2.5				

## 7. NOTES ON NIST TIME SCALES AND PRIMARY STANDARDS

The frequencies of the time scales, TA(NIST) and UTC(NIST), are calibrated with the NIST primary frequency standards. The UTC(NIST) scale is coordinated within a microsecond of the internationally coordinated time scale, UTC, generated at the BIH. It is used to control all of the NIST time and frequency services. The last calibration of the relative frequency offset,  $y$ , of UTC(NIST) as generated in Boulder, Colorado, gave:

$$1) \quad y_{UTC(NIST)}(\text{July 1987}) - y_{NBS-6}(\text{July 1987}) = (-0.6 \pm 2 (1 \text{ sigma})) \times 10^{-13}$$

for the date shown. This calibration includes a correction for the systematic offset due to room temperature blackbody radiation, which is approximately  $(\Delta y_{BB}) = -1.7 \times 10^{-14}$ . Using GPS<sup>1</sup>, the frequency of TAI for the dates shown were measured to be:

$$2) \quad y_{TAI}(\text{July 1987}) - y_{NBS-6}(\text{July 1987 on geoid}) = (+1.7 \pm 2 (1 \text{ sigma})) \times 10^{-13}$$

where account has been taken of the gravitational "red shift."

Starting 1 January 1975, an accuracy algorithm was implemented to bring the second used in the generation of TA(NIST) closer to the NIST "best estimate" of the SI second (see references, p.6). The relative frequency associated with this "best estimate" is denoted  $y_{Cs(NIST)}$ . The last calibration (July 1987) covered the period from October 1986 through July 1987.

$$3) \quad y_{Cs(NIST)} - y_{NBS-6} = (+1.4 \pm 2) \times 10^{-13} \text{ (July 1987)}$$

and

$$4) \quad y_{TAI} - y_{Cs(NIST)} \text{ on geoid} = (+0.3 \pm 0.7) \times 10^{-13} \text{ (July 1987)}$$

This algorithm should provide nearly optimum accuracy and stability for TA(NIST) since it uses all past frequency calibrations with the NIST primary standards. These calibrations are weighted proportionately to the frequency memory of the clock ensemble that generates atomic time. This algorithm, therefore, capitalizes on a weighted combination of all the frequency calibrations with the primary standards in order to gain a "best estimate" of the SI Second while simultaneously obtaining the best uniformity available from the ensemble of working clocks in the atomic time scale system. The relative frequency of TA(NIST) is steered toward  $y_{Cs(NIST)}$  by slight frequency drift corrections of the order of 1 part in  $10^{13}/\text{yr}$ .

TA(NIST) and UTC(NIST) are no longer simply related by an equation. TA(NIST) is now computed each month using a Kalman algorithm which minimizes the mean square time dispersion. UTC(NIST) is now independently computed using a different algorithm and is steered in frequency to keep its time within a microsecond of UTC(BIH). Table 7.1 lists monthly values of the time difference between UTC(NIST) and TA(NIST). A linear interpolation between monthly values will typically be within 10 ns of the actual time difference, TA(NIST) - UTC(NIST).

The primary standards of NIST (NBS-4 and NBS-6) are used in either of two modes: as calibrators of frequency to provide a reference for the SI second; or as member clocks of the NIST clock ensemble, to help keep the proper time for TA(NIST) and the coordinated time for UTC(NIST). Operating in the clock mode, NBS-4 and NBS-6 are only used and weighted according to their stability performance. Accuracy enters only when they are used as frequency calibrators, in which case clock operation is necessarily interrupted.

---

<sup>1</sup>GPS is the Global Positioning System, a network of navigation satellites.

Table 7.1 is a list of changes in the time scale frequencies of both TA(NIST) and UTC(NIST) as well as a list of the time and frequency differences between TA(NIST) and UTC(NIST) at the dates of leap seconds, and/or frequency or frequency drift changes.

TABLE 7.1

FREQUENCY CHANGES					
DATE	(MJD)	TA(NIST)	UTC(NIST)	TA(NIST) - UTC(NIST)	y[UTC(NIST)] - y[TA(NIST)]
1 Apr 88	47252	0	-1.50 ns/d	24.045 108 130 s	-3.85 E-13
1 May 88	47282	0	-1.50 ns/d	24.045 109 170 s	-4.29 E-13
1 Jun 88	47313	0	-1.50 ns/d	24.045 110 358 s	-4.47 E-13
1 Jul 88	47343	0	-1.60 ns/d	24.045 111 523 s	-4.64 E-13
1 Aug 88	47374	0	-0.40 ns/d	24.045 112 802 s	-4.89 E-13
1 Sep 88	47405	0	-1.00 ns/d	24.045 114 144 s	-5.15 E-13
1 Oct 88	47435	0	1.00 ns/d	24.045 114 515 s	-5.15 E-13
1 Nov 88	47466	0	1.25 ns/d	24.045 116 854 s	-4.88 E-13
1 Dec 88	47496	0	1.50 ns/d	24.045 118 088 s	-4.69 E-13
1 Jan 89	47527	0	1.50 ns/d	24.045 119 325 s	-4.57 E-13
1 Feb 89	47558	0	1.00 ns/d	24.045 120 538 s	-4.51 E-13
1 Mar 89	47586	0	-1.25 ns/d	24.045 121 622 s	-4.58 E-13
1 Apr 89	47617	0	-1.50 ns/d	24.045 122 871 s	-4.66 E-13
1 May 89	47647	0	-1.50 ns/d	24.045 124 078 s	-4.75 E-13
1 Jun 89	47678	0	-1.00 ns/d	24.045 125 375 s	-4.92 E-13
1 Jul 89	47708	0	-1.00 ns/d	24.045 126 670 s	-5.09 E-13
1 Aug 89	47739	0	-1.00 ns/d	24.045 128 060 s	-5.35 E-13
1 Sep 89	47770	0	-1.00 ns/d	24.045 129 538 s	-5.58 E-13

UTC(NIST) is steered in time toward UTC by occasional frequency changes of the order of a few nanoseconds per day; 1 ns/d is approximately  $1.16E-14$ . Otherwise, y[UTC(NIST)] is maintained as stable as possible.

#### REFERENCES

- Allan, David W., et al., "An accuracy algorithm for an atomic time scale," Metrologia, Vol.11, No.3, pp.133-138 (September 1975).
- Glaze, D.J., et al., "NBS-4 and NBS-6: The NIST primary frequency standards," Metrologia, Vol.13, pp.17-28 (1977).
- Wineland, D.J., et al., "Results on limitations in primary cesium standard operation," IEEE Trans. on Instr. and Meas., Vol.IM-25, No.4, pp.453-458 (December 1976).
- Allan, David W. and Weiss, Marc, "Accurate Time and Frequency Transfer During Common View of a GPS Satellite," Proc. 34th Annual Symposium on Frequency Control, p.334 (1980).
- Allan, David W. and Barnes, James A., "Optimal Time & Frequency using GPS signals," Proc. 36th Annual Symposium on Frequency Control, p.378 (1982).

## 8. SPECIAL ANNOUNCEMENTS

### AUTOMATED COMPUTER TIME SERVICE (ACTS)

On March 9, 1988, NIST initiated operation of a telephone time service designed to provide computers with telephone access to NIST time at accuracies approaching 1 ms. Features of the service include automated compensation for telephone-line delay, advanced alert for changes to and from daylight savings time and advanced notice of insertion of leap seconds. The ASCII-character time code should operate with standard modems and most computer systems. While the system can be used to set computer time-of-day clocks, simple hardware can also be developed to set other clock systems.

The test phase for this service is now complete and NIST is committed to long-term operation of the service. Additional lines will be added as use expands. NIST requests that calling times be spread out so that the system is not heavily taxed in some narrow time frame (e.g., midnight). The service telephone number is (303) 494-4774. The number may be changed at a later date. A help message can be obtained by returning a ? during the first 6 s of transmission.

With appropriate user software, the NIST-ACTS service provides three modes for checking and/or setting computer time-of-day clocks.

1. In the simplest form of the (1200 Baud) service, the user receives the time code and an on-time marker/character which has been advanced a fixed period to nominally account for modem and telephone-line delays. Accuracy in this mode should be no worse than 0.1 s unless the connection is routed through a satellite.

2. At 1200 Baud, if the user's system echoes all characters to NIST, the round-trip line delay will be measured and the on-time marker advanced to compensate for that delay. The accuracy in this mode should be better than 10 ms. Our experience to date indicates that the asymmetry in conventional, 1200-Baud modems limits the accuracy at this level. Repeatability is about 1 ms.

3. At 300 Baud the user can obtain the same type of service as described in item 2 above, but there is generally less problem with modem asymmetry at this rate and our experience indicates that the accuracy is about 1 ms.

The accuracy statements here are based upon the assumption that the telephone connection is reciprocal, that is, that both directions of communication follow the same path with the same delay. Discussions with telephone carriers indicate that this is the general mode of operation and our tests to date indicate that the lines are both stable and reciprocal.

In order to assist users of the service, NIST has developed documentation of the features of the service, some example software which can be used in conjunction with certain popular personal computers and simple circuitry which can be used to extract an on-time pulse. This material is available on a 5 $\frac{1}{4}$ -in, 360-kbyte DOS diskette with instructions for \$35.00 from the NIST Office of Standard Reference Materials, B311-Chemistry Bldg, NIST, Gaithersburg, MD, 20899, (301) 975-6776. Specify the Automated Computer Time Service, RM8101. Further technical questions and comments should be directed to NIST-ACTS, NIST Time and Frequency Division, 325 Broadway, Boulder, CO 80303.

### 44TH ANNUAL FREQUENCY CONTROL SYMPOSIUM

The 44th Annual Frequency Control Symposium will be held May 23 - 25, 1990 in Baltimore, MD. This symposium is the leading technical conference addressing all aspects of frequency control and precision timekeeping. Authors are invited to submit papers dealing with recent progress in research, development and applications in areas represented by the following topics:

- Fundamental properties of piezoelectric crystals
- Theory and design of piezoelectric resonators
- Resonator processing techniques
- Filters
- Surface acoustic wave devices (SAW)
- Quartz crystal oscillators
- Microwave and millimeter wave oscillators
- Signal processing and frequency control circuitry
- Atomic and molecular frequency standards
- Frequency and time coordination and distribution
- Sensors and transducers
- Applications of frequency control
- Measurement and specifications

Contact: Dr. R. L. Filler, US Army Electronics Technology and Devices Laboratory, ATTN: SLGET-EQ, Fort Monmouth, NJ 07703-5000; (201) 544-2467.

U.S. DEPT. OF COMMERCE <b>BIBLIOGRAPHIC DATA SHEET</b> <i>(See instructions)</i>	1. PUBLICATION OR REPORT NO. NISTIR 89-3910-10	2. Performing Agency Report No. B89-0064	3. Publication Date October 1989
4. TITLE AND SUBTITLE <p style="text-align: center;"><b>Time &amp; Frequency Bulletin</b></p>			
5. AUTHOR(S)			
6. PERFORMING ORGANIZATION (If joint or other than NBS, see instructions) NATIONAL BUREAU OF STANDARDS DEPARTMENT OF COMMERCE WASHINGTON, D.C. 20234		7. Contract/Grant No.	8. Type of Report & Period Covered
9. SPONSORING ORGANIZATION NAME AND COMPLETE ADDRESS (Street, City, State, ZIP)			
10. SUPPLEMENTARY NOTES  <input type="checkbox"/> Document describes a computer program; SF-185, FIPS Software Summary, is attached.			
11. ABSTRACT (A 200-word or less factual summary of most significant information. If document includes a significant bibliography or literature survey, mention it here)  <p style="text-align: center;">The Time &amp; Frequency Bulletin provides information on performance of time scales and a variety of broadcasts (and related information) to users of the NIST services.</p>			
12. KEY WORDS (Six to twelve entries; alphabetical order; capitalize only proper names; and separate key words by semicolons) <p style="text-align: center;">clocks; dissemination; frequency; GPS; oscillators; time</p>			
13. AVAILABILITY <input type="checkbox"/> Unlimited <input checked="" type="checkbox"/> For Official Distribution. Do Not Release to NTIS <input type="checkbox"/> Order From Superintendent of Documents, U.S. Government Printing Office, Washington, D.C. 20402.  <input type="checkbox"/> Order From National Technical Information Service (NTIS), Springfield, VA. 22161			14. NO. OF PRINTED PAGES <p style="text-align: center;">8</p> 15. Price





AN EQUAL OPPORTUNITY EMPLOYER

U.S. Department of Commerce  
National Institute of Standards and Technology  
Time & Frequency Division  
Boulder, Colorado 80303

---

FIRST-CLASS MAIL  
POSTAGE & FEES PAID  
NIST  
Permit No. G195

OFFICIAL BUSINESS  
Penalty for Private Use, \$300

---

GWEN BENNETT  
DIVISION 524 COPY 2  
NATL BUREAU OF STANDARDS  
325 BROADWAY  
BOULDER, CO 80303